

保康潔 Coversafe™ 不受清潔劑影響抗菌效能

Type of Cleaning Product 清潔劑種類	Efficiency 抗菌活性率 ^{#3}
None (Control) 不含清潔劑標本	>99.999% (R=5.86)
Isopropyl alcohol 異丙醇	>99.999% (R=5.86)
Surfanios Premium (hospital disinfectant) 醫院消毒劑	>99.999% (R=5.86)
Cleaning agent with bleach 漂白水清洗劑	>99.999% (R=5.86)

#3：每個測試樣本先接觸 100 次清潔劑，然後再以 JIS Z 2801:2010 方法檢測抗大腸桿菌活性，
測試機構：[FONDEREPHAR](#)

證書編號：20-2663



Reference 20-2663

CERTIFICATE OF ANALYSIS

Society : PYLOTE
Address : 22 Avenue de la Mouyssaguèse
31280 Dremil-Lafage
FRANCE

To the attention of : LOIC MARCHIN

Customer Reference:	Coversafe Film (Film Gerg. ADD)
Fonderephar Samples Reference:	20-2663-2/20-2663-3/20-2663-4/20-2663-5
Date of sample receipt:	June 2 nd , 2020
Date of sample analysis:	June 2020
Date of certificate of analysis:	June 17 th , 2020

Test
Evaluation of antimicrobial efficiency based on JIS Z2801 : 2010 for bacteria

Results: The results are given as log reduction R, corresponding to the value of antimicrobial activity after a contact time 24H

Escherichia coli CIP 53.126

- R = 5,86 without wash
- R = 5,86 after 100 washes with isopropilic alcohol
- R = 5,86 after 100 washes with detergent with bleach
- R = 5,86 after 100 washes with Surfianis Premium

Certified by Catherine FEUILLOLAY
Test Manager